



PURGED

Food and Drug Administration
Minneapolis District
240 Hennepin Avenue
Minneapolis MN 55401-1999
Telephone: 612-334-4100

November 3, 2000

WARNING LETTER

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

Refer to MIN 01-10

Carol Andrews
Facility Administrator
Aurora Medical Group
6745 West Wells Street
Wauwatosa, Wisconsin 53216

Dear Ms. Andrews:

On October 26, 2000, a representative of the State of Wisconsin, acting on behalf of the Food and Drug Administration (FDA), inspected your facility (FDA Certificate #185884). This inspection revealed a serious regulatory problem involving the mammography at your facility.

Under a United States Federal law, the Mammography Quality Standards Act of 1992 (MQSA), your facility must meet specific requirements for mammography. These requirements help protect the health of women by assuring that a facility can perform quality mammography. Based on the documentation your site presented at the time of the inspection, the following Level 2 findings were documented at your facility:

Repeat Level 2 Non-Compliance:

1. Corrective action before further exams, for a failing image score, or a phantom background optical density, or density difference outside the allowable regulatory limits, was not documented for mammography unit #1 (~~~~~).

Level 2 Non-Compliance:

2. Film processor QC records were missing four out of 19 days of operation in June 2000 (21%). Mammography film processor = ~~~~~ or

Page Two

Carol Andrews

November 3, 2000

~~~~~; Room = darkroom. Processor QC is mandatory for each day clinical images are developed.

3. Film processor QC records were missing three consecutive days of operation. Mammography film processor = ~~~~~; Room = darkroom.

The specific problems noted above appeared on your MQSA Facility Inspection Report which was issued to your facility following the close of the inspection.

Because these conditions may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, they represent a serious violation of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction, charging your facility for the cost of on-site monitoring, assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, the Standards, suspension or revocation of your facility's FDA certificate, or obtaining a court injunction against further mammography.

It is necessary for you to act on this matter immediately. Please explain to this office in writing within 15 working days from the date you received this letter:

- The specific steps you have taken to correct all of the violations noted in this letter;
- Each step your facility is taking to prevent the recurrence of similar violations;
- Equipment settings (including technique factors), raw test data, and calculated final results, where appropriate; and
- Sample records that demonstrate proper record keeping procedures if the findings relate to quality control or other records.

Please submit your response to Thomas W. Garvin, Radiological Health Specialist, Food and Drug Administration 2675 N. Mayfair Road, Suite 200, Milwaukee, WI 53226-1305.

Finally, you should understand that there are many FDA requirements pertaining to mammography. This letter pertains only to findings of your inspection and does not necessarily address other obligations you have under the law. You may obtain general information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug Administration, P.O. Box 6057, Columbia, MD 21045-6057 (1-800-838-7715) or through the Internet at <http://www.fda.gov/cdrh/mammography/index.html>.

Page Three

Carol Andrews  
November 3, 2000

If you have specific questions about mammography facility requirements or about the content of this letter please feel free to phone Mr. Garvin at (414) 771-7167 ext. 12.


Sincerely,



James A. Rahto  
Director  
Minneapolis District

TWG/ccl

*Sc*

xc:   
Lead Interpreting Radiologist  
Aurora Medical Group  
6745 West Wells St.  
Wauwatosa, WI 53216

G. Edwin Howe  
President  
Aurora Health Care, Inc.  
3030 West Montana Ave.  
Milwaukee, WI 53215

Paul Schmidt  
Chief, Radiation Protection Unit  
State of Wisconsin  
P.O. Box 2659  
Madison, WI 53701-2659

Priscilla F. Butler  
Director, Breast Imaging Accreditation Programs  
American College of Radiology  
1891 Preston White Drive  
Reston, VA 20191